

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4981

Name: RITTER 1  
Sterilizer ID: V2178198  
10/09/2024 01:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.6 F 0.0  
2:00 124.7 F 0.1  
4:00 133.3 F 0.7  
6:00 157.3 F 0.9  
8:00 192.9 F 3.2  
10:00 210.1 F 6.8  
12:00 222.4 F 10.6  
14:00 231.0 F 7.2  
16:00 233.9 F 10.2  
18:00 239.5 F 8.9  
20:00 243.3 F 12.7  
22:00 241.4 F 11.1  
24:00 253.6 F 17.2  
26:00 264.2 F 23.4  
28:00 271.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 271.3 F 28.5  
2:00 272.3 F 29.0  
3:00 271.9 F 28.6  
4:00 271.1 F 28.2

Min 271.0 F 28.1  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 209.1 F 0.0  
10:00 199.7 F 0.0  
15:00 212.4 F 0.0  
20:00 231.4 F 0.0  
25:00 233.0 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 28:03  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:06:38  
CYCLE COMPLETE

10/09/2024 02:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_