
Midmark M11 - v1.0.0
Total Cycles: 1722

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.3 F 0.0
2:00 175.4 F 0.7
4:00 173.8 F 2.4
6:00 191.6 F 5.4
8:00 208.5 F 9.2
10:00 223.9 F 13.5
12:00 227.4 F 8.3
14:00 239.5 F 13.0
16:00 239.6 F 10.2
18:00 236.1 F 8.3
20:00 249.0 F 14.4
22:00 260.5 F 20.8
24:00 270.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.5 F 28.8
2:00 272.6 F 28.6
3:00 272.5 F 28.6
4:00 272.5 F 28.6

Min 270.7 F 27.8
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 232.7 F 0.0
10:00 245.9 F 0.0
15:00 234.3 F 0.0
20:00 226.7 F 0.0
25:00 222.0 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

03/16/2022 11:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____