

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1654

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2022 12:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
2:00 197.6 F 3.4  
4:00 197.8 F 8.8  
6:00 226.2 F 9.1  
8:00 233.4 F 10.6  
10:00 234.5 F 8.2  
12:00 241.4 F 10.6  
14:00 246.0 F 13.9  
16:00 258.5 F 20.8  
18:00 269.7 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.2  
1:00 272.2 F 30.2  
2:00 272.0 F 29.7  
3:00 272.9 F 30.1  
4:00 272.4 F 30.4  
  
Min 271.0 F 29.1  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 211.3 F 0.0  
10:00 223.8 F 0.0  
15:00 224.0 F 0.0  
20:00 222.4 F 0.0  
25:00 221.3 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 2:42  
HEATING: 19:00  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

10/25/2022 01:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_