
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001257
11:31 15/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:25
STERILIZING 5:44
136.0 C 220kPa 5:44
136.2 C 223kPa 6:14
136.3 C 224kPa 6:44
136.4 C 224kPa 7:14
136.4 C 226kPa 7:44
136.2 C 223kPa 8:14
136.4 C 224kPa 8:44
136.2 C 223kPa 9:14
136.2 C 223kPa 9:14
VENTING 9:14
AIR DRYING 10:13
CYCLE COMPLETE 25:13

Digital Signature #
B4C8CB8FBB5A62A3
