
Midmark M11 - v1.0.5
Total Cycles: 2709

Name: RITTER2
Sterilizer ID: V2313228
07/27/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 169.8 F 1.0
4:00 186.0 F 5.1
6:00 204.9 F 10.4
8:00 222.9 F 6.0
10:00 234.9 F 11.6
12:00 234.3 F 9.0
14:00 238.7 F 10.0
16:00 246.9 F 14.2
18:00 259.0 F 20.8
20:00 269.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 28.9
3:00 272.4 F 29.9
4:00 272.2 F 29.4

Min 271.2 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 211.9 F 0.0
10:00 220.7 F 0.0
15:00 224.8 F 0.0
20:00 223.5 F 0.0
25:00 222.1 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

07/27/2023 12:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____