
Midmark M11 - v1.0.0
Total Cycles: 1869

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.5 F 0.0
2:00 122.9 F 0.1
4:00 135.5 F 1.5
6:00 164.6 F 3.7
8:00 191.0 F 7.2
10:00 207.9 F 11.4
12:00 219.6 F 4.5
14:00 229.8 F 9.2
16:00 236.2 F 10.6
18:00 240.4 F 11.0
20:00 236.9 F 8.8
22:00 250.7 F 15.4
24:00 262.2 F 22.0
26:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.7
2:00 271.5 F 28.2
3:00 272.4 F 28.7
4:00 271.5 F 28.0

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 202.9 F 0.0
10:00 201.1 F 0.0
15:00 203.1 F 0.0
20:00 205.2 F 0.0
25:00 231.3 F 0.0
30:00 246.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:35
CYCLE COMPLETE

04/26/2022 10:20 AM
ACCEPTED: X
REJECTED:
Operator: U02