

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1542

Name: RITTER 1  
Sterilizer ID: V2178198  
01/20/2022 07:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.2 F 0.0  
2:00 77.4 F 0.0  
4:00 81.6 F 0.0  
6:00 107.5 F 0.9  
8:00 138.6 F 3.5  
10:00 166.8 F 6.7  
12:00 196.0 F 10.9  
14:00 208.2 F 3.4  
16:00 215.7 F 6.6  
18:00 226.4 F 10.4  
20:00 232.6 F 8.3  
22:00 239.2 F 11.1  
24:00 235.3 F 8.3  
26:00 248.7 F 14.5  
28:00 259.5 F 20.5  
30:00 269.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.3 F 28.8  
2:00 271.6 F 28.4  
3:00 272.3 F 28.7  
4:00 271.5 F 28.4  
  
Min 271.4 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 210.3 F 0.0  
10:00 218.3 F 0.0  
15:00 213.0 F 0.0  
20:00 216.2 F 0.0  
25:00 221.7 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:02  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:09  
CYCLE COMPLETE

01/20/2022 08:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_