
Midmark M11 - v1.0.5
Total Cycles: 1468

Name: RITTER2
Sterilizer ID: V2313228
09/08/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 198.9 F 2.5
4:00 194.5 F 7.7
6:00 211.5 F 13.3
8:00 225.7 F 8.6
10:00 237.6 F 11.1
12:00 240.9 F 12.1
14:00 240.2 F 11.0
16:00 253.6 F 17.8
18:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.6 F 29.0
3:00 272.2 F 29.9
4:00 272.2 F 29.5

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 205.0 F 0.0
10:00 225.4 F 0.0
15:00 220.1 F 0.0
20:00 221.8 F 0.0
25:00 218.9 F 0.0
30:00 224.6 F 0.0

Drying Complete

FILLING: 1:53
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

09/08/2022 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____