
Midmark M11 - v1.0.5
Total Cycles: 3190

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 165.3 F 2.3
4:00 181.5 F 6.3
6:00 201.5 F 10.6
8:00 219.9 F 6.6
10:00 224.2 F 10.1
12:00 234.3 F 8.6
14:00 242.1 F 13.1
16:00 242.1 F 12.0
18:00 254.6 F 18.4
20:00 265.4 F 25.1
22:00 271.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.5 F 29.7
2:00 272.9 F 30.0
3:00 272.6 F 30.3
4:00 271.9 F 29.9

Min 271.8 F 29.7
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 236.1 F 0.0
10:00 230.7 F 0.0
15:00 219.6 F 0.0
20:00 212.9 F 0.0
25:00 211.6 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

12/06/2023 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____