
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000915
16:37 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:34
136.0 C 228kPa 4:34
136.2 C 231kPa 5:04
136.4 C 233kPa 5:34
136.3 C 232kPa 6:04
136.4 C 233kPa 6:34
136.1 C 231kPa 7:04
136.4 C 233kPa 7:34
136.4 C 233kPa 8:04
136.4 C 233kPa 8:04
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 20:45

Digital Signature #
9B16D474500B6651
