
Midmark M11 - v1.0.0
Total Cycles: 1124

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2021 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.2 F 0.0
2:00 132.8 F 0.9
4:00 159.7 F 3.7
6:00 187.4 F 7.9
8:00 205.6 F 13.5
10:00 220.9 F 7.0
12:00 235.1 F 12.2
14:00 234.0 F 8.8
16:00 240.0 F 9.8
18:00 245.9 F 13.2
20:00 258.5 F 20.0
22:00 269.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.4 F 28.3
3:00 271.8 F 29.0
4:00 272.3 F 28.9

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 232.7 F 0.0
10:00 218.7 F 0.0
15:00 219.3 F 0.0
20:00 209.6 F 0.0
25:00 190.2 F 0.0
Dry Cycle Aborted

FILLING: 1:13
HEATING: 22:57
STERILIZING: 4:00
VENTING: 2:54
DRYING: 26:59
TOTAL CYCLE: 00:58:03

08/27/2021 11:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____