
Midmark M11 - v1.0.0
Total Cycles: 1880

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 191.6 F 3.8
4:00 192.1 F 8.4
6:00 215.0 F 12.3
8:00 224.5 F 8.1
10:00 238.4 F 13.4
12:00 237.9 F 10.5
14:00 237.2 F 8.5
16:00 250.6 F 15.5
18:00 262.2 F 22.4
20:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.5 F 28.3
3:00 272.6 F 29.3
4:00 272.1 F 28.7

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 219.6 F 0.0
10:00 228.6 F 0.0
15:00 230.8 F 0.0
20:00 229.2 F 0.0
25:00 228.2 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

04/27/2022 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____