
Midmark M11 - v1.0.5
Total Cycles: 612

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 181.8 F 3.5
4:00 198.7 F 8.9
6:00 217.9 F 8.6
8:00 230.5 F 10.4
10:00 231.9 F 7.8
12:00 240.6 F 11.1
14:00 246.5 F 14.0
16:00 259.3 F 21.0
18:00 270.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.5
2:00 271.7 F 29.3
3:00 272.3 F 29.4
4:00 271.5 F 29.1

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 231.5 F 0.0
10:00 221.7 F 0.0
15:00 228.7 F 0.0
20:00 220.9 F 0.0
25:00 218.6 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

12/08/2021 01:07 PM
ACCEPTED: X
REJECTED:
Operator: U02