

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2605

Name: RITTER2  
Sterilizer ID: V2313228  
07/04/2023 04:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
2:00 194.6 F 4.1  
4:00 209.0 F 9.9  
6:00 223.2 F 6.7  
8:00 235.5 F 11.7  
10:00 236.3 F 9.8  
12:00 236.6 F 8.9  
14:00 250.5 F 16.0  
16:00 262.8 F 23.1  
18:00 272.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.3 F 29.4  
2:00 271.7 F 29.3  
3:00 272.3 F 29.3  
4:00 271.5 F 29.2  
  
Min 271.4 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 213.6 F 0.0  
10:00 220.2 F 0.0  
15:00 220.3 F 0.0  
20:00 218.3 F 0.0  
25:00 216.5 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 2:55  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

07/04/2023 04:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_