
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002355
9:50 05/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:28
STERILIZING 3:29
136.0 C 220kPa 3:29
136.3 C 223kPa 3:59
136.4 C 225kPa 4:29
136.4 C 225kPa 4:59
136.4 C 225kPa 5:29
136.2 C 220kPa 5:59
136.4 C 221kPa 6:29
136.2 C 222kPa 6:59
136.2 C 222kPa 6:59
VENTING 6:59
AIR DRYING 7:51
CYCLE COMPLETE 7:56

Digital Signature #
208B31E6C2C65E7B
