
Midmark M11 - v1.0.0
Total Cycles: 4444

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2024 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 134.8 F 1.5
4:00 166.8 F 4.2
6:00 168.5 F 7.9
8:00 189.6 F 11.8
10:00 211.3 F 3.9
12:00 219.6 F 8.0
14:00 231.3 F 12.2
16:00 232.7 F 8.6
18:00 243.3 F 11.9
20:00 243.0 F 12.2
22:00 254.5 F 18.0
24:00 264.6 F 24.1
26:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.2
2:00 272.5 F 29.6
3:00 271.7 F 28.9
4:00 272.4 F 29.4

Min 271.6 F 28.9
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 224.1 F 0.0
10:00 218.1 F 0.0
15:00 222.6 F 0.0
20:00 225.3 F 0.0
25:00 232.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:45
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:34
CYCLE COMPLETE

04/10/2024 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____