
Midmark M11 - v1.0.0
Total Cycles: 5257

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 87.6 F 0.0
4:00 93.6 F 0.1
6:00 121.4 F 2.0
8:00 151.5 F 4.9
10:00 186.7 F 8.6
12:00 186.7 F 12.3
14:00 211.2 F 3.6
16:00 223.2 F 6.8
18:00 232.5 F 10.4
20:00 230.4 F 8.2
22:00 238.7 F 10.5
24:00 236.4 F 8.1
26:00 247.9 F 14.0
28:00 259.4 F 20.3
30:00 269.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.7 F 28.8
2:00 271.8 F 28.3
3:00 272.0 F 28.2
4:00 272.8 F 28.9

Min 271.6 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 204.2 F 0.0
10:00 217.5 F 0.0
15:00 205.9 F 0.0
20:00 216.8 F 0.0
25:00 220.1 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

01/22/2025 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____