
Midmark M11 - v1.0.0
Total Cycles: 4074

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2023 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.5 F 0.0
2:00 138.2 F 1.5
4:00 146.4 F 4.6
6:00 171.8 F 8.1
8:00 198.8 F 12.3
10:00 209.9 F 4.7
12:00 221.6 F 9.1
14:00 235.5 F 11.6
16:00 237.1 F 10.3
18:00 236.2 F 8.6
20:00 248.0 F 14.4
22:00 259.4 F 20.6
24:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.2
2:00 272.0 F 28.7
3:00 271.2 F 28.5
4:00 272.4 F 29.2

Min 271.2 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 228.3 F 0.0
10:00 227.5 F 0.0
15:00 230.9 F 0.0
20:00 225.4 F 0.0
25:00 229.0 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

12/01/2023 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____