
Midmark M11 - v1.0.5
Total Cycles: 285

Name: RITTER2
Sterilizer ID: V2313228
08/18/2021 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 171.6 F 2.9
4:00 186.6 F 7.4
6:00 207.6 F 12.4
8:00 220.8 F 6.7
10:00 235.2 F 12.1
12:00 238.5 F 10.6
14:00 239.1 F 10.3
16:00 253.2 F 17.2
18:00 264.9 F 24.2
20:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.1 F 29.7
2:00 272.4 F 29.8
3:00 272.5 F 29.8
4:00 272.4 F 29.8

Min 271.7 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 234.3 F 0.0
10:00 235.0 F 0.0
15:00 223.5 F 0.0
20:00 217.9 F 0.0
25:00 215.6 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

08/19/2021 06:40 AM
ACCEPTED: X
REJECTED:
Operator: U02