
Midmark M11 - v1.0.0
Total Cycles: 3450

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.6 F 0.0
2:00 171.4 F 1.2
4:00 177.6 F 4.3
6:00 197.6 F 8.0
8:00 214.2 F 12.8
10:00 223.6 F 6.7
12:00 236.0 F 11.4
14:00 233.6 F 8.2
16:00 243.6 F 12.2
18:00 243.9 F 12.4
20:00 255.9 F 18.6
22:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.3
2:00 272.3 F 28.8
3:00 271.6 F 28.4
4:00 272.5 F 29.2

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 218.7 F 0.0
10:00 221.8 F 0.0
15:00 223.7 F 0.0
20:00 226.2 F 0.0
25:00 226.1 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

06/27/2023 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____