
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002128
8:08 07/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:33
STERILIZING 4:45
136.0 C 220kPa 4:45
136.4 C 223kPa 5:15
136.0 C 221kPa 5:45
136.3 C 223kPa 6:15
136.3 C 224kPa 6:45
136.1 C 222kPa 7:15
136.4 C 224kPa 7:45
136.1 C 222kPa 8:15
136.1 C 222kPa 8:15
VENTING 8:15
AIR DRYING 9:09
CYCLE COMPLETE 24:25

Digital Signature #
3EEB0267856DB310
