

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2154

Name: RITTER2  
Sterilizer ID: V2313228  
03/16/2023 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.5 F 0.0  
2:00 143.6 F 1.7  
4:00 154.0 F 4.5  
6:00 178.2 F 8.6  
8:00 201.3 F 10.1  
10:00 221.2 F 7.7  
12:00 232.7 F 12.9  
14:00 236.0 F 10.1  
16:00 235.9 F 8.6  
18:00 248.3 F 15.0  
20:00 260.3 F 21.8  
22:00 271.2 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.2  
1:00 272.1 F 29.9  
2:00 272.2 F 29.5  
3:00 271.3 F 29.3  
4:00 272.6 F 30.0

Min 271.3 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
5:00 230.7 F 0.0  
10:00 225.3 F 0.0  
15:00 221.8 F 0.0  
20:00 218.1 F 0.0  
25:00 217.6 F 0.0  
30:00 216.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:33  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

03/16/2023 10:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_