
Midmark M11 - v1.0.0
Total Cycles: 2966

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2023 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.5 F 0.0
2:00 185.2 F 0.1
4:00 197.4 F 0.5
6:00 210.7 F 4.4
8:00 218.3 F 9.1
10:00 231.3 F 10.1
12:00 235.1 F 10.4
14:00 233.3 F 7.8
16:00 243.2 F 11.8
18:00 245.4 F 13.0
20:00 258.3 F 19.9
22:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.4
3:00 271.9 F 29.0
4:00 272.5 F 29.0

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 224.9 F 0.0
10:00 235.0 F 0.0
15:00 240.5 F 0.0
20:00 227.5 F 0.0
25:00 229.3 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:10
CYCLE COMPLETE

03/02/2023 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____