
Midmark M11 - v1.0.0
Total Cycles: 1734

Name: RITTER 1
Sterilizer ID: V2178198
03/23/2022 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 190.7 F 2.8
4:00 194.7 F 7.1
6:00 214.6 F 12.8
8:00 227.0 F 7.5
10:00 239.7 F 12.7
12:00 238.5 F 10.3
14:00 236.1 F 8.3
16:00 250.5 F 15.3
18:00 262.2 F 22.1
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.0 F 28.3
2:00 272.4 F 28.7
3:00 272.3 F 29.0
4:00 271.7 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 231.0 F 0.0
10:00 230.9 F 0.0
15:00 229.0 F 0.0
20:00 228.9 F 0.0
25:00 227.2 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

03/23/2022 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____