
Midmark M11 - v1.0.0
Total Cycles: 1228

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2021 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.5 F 0.0
2:00 151.3 F 1.9
4:00 169.0 F 4.8
6:00 190.3 F 8.3
8:00 211.0 F 12.7
10:00 224.3 F 7.0
12:00 237.0 F 11.9
14:00 234.9 F 8.4
16:00 242.1 F 11.6
18:00 244.8 F 12.5
20:00 256.5 F 18.6
22:00 267.0 F 25.1
24:00 272.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.8
2:00 272.1 F 28.9
3:00 272.3 F 28.5
4:00 272.5 F 29.1

Min 271.5 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 236.4 F 0.0
10:00 223.2 F 0.0
15:00 221.1 F 0.0
20:00 214.4 F 0.0
25:00 180.0 F 0.0
30:00 168.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

10/04/2021 11:37 AM
ACCEPTED: X
REJECTED:
Operator: U01