
Midmark M11 - v1.0.0
Total Cycles: 725

Name: RITTER 1
Sterilizer ID: V2178198
03/17/2021 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 128.9 F 1.0
4:00 136.2 F 3.2
6:00 162.3 F 6.4
8:00 189.3 F 10.5
10:00 218.8 F 4.9
12:00 221.5 F 8.3
14:00 233.9 F 12.9
16:00 235.3 F 10.0
18:00 235.8 F 8.6
20:00 247.8 F 14.4
22:00 259.0 F 20.8
24:00 270.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.4
2:00 271.4 F 29.0
3:00 272.1 F 29.4
4:00 271.2 F 28.8

Min 271.2 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 205.4 F 0.0
10:00 220.8 F 0.0
15:00 232.3 F 0.0
20:00 194.2 F 0.0
25:00 191.3 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

03/17/2021 01:10 PM
ACCEPTED: X
REJECTED:
Operator: U01