

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3293

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2024 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.7 F 0.0  
2:00 137.6 F 0.3  
4:00 140.3 F 2.3  
6:00 164.6 F 5.7  
8:00 185.4 F 9.9  
10:00 202.5 F 7.7  
12:00 220.4 F 7.5  
14:00 232.3 F 12.0  
16:00 232.0 F 8.9  
18:00 238.8 F 10.8  
20:00 244.9 F 13.4  
22:00 256.7 F 19.7  
24:00 267.4 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.7 F 30.0  
2:00 271.9 F 29.4  
3:00 271.4 F 29.3  
4:00 272.5 F 30.1

Min 271.3 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 217.1 F 0.0  
10:00 217.3 F 0.0  
15:00 217.8 F 0.0  
20:00 218.6 F 0.0  
25:00 236.5 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:29  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:03:58  
CYCLE COMPLETE

01/12/2024 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_