

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001215  
9:46 19/02/2021  
RUBBER AND PLASTICS  
121 C for 15 min.  
CYCLE START 0:00  
CYCLE ABORTED 0:02  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 0:02  
-----

Digital Signature #  
40C9AC36C8477A1C  
-----