
Midmark M11 - v1.0.0
Total Cycles: 2711

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.4 F 0.0
2:00 205.7 F 4.3
4:00 209.0 F 9.0
6:00 223.7 F 6.7
8:00 234.5 F 10.4
10:00 236.8 F 7.5
12:00 242.7 F 11.3
14:00 246.8 F 13.5
16:00 259.9 F 20.7
18:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.7 F 28.9
4:00 271.7 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.3 F 0.0
5:00 228.0 F 0.0
10:00 236.2 F 0.0
15:00 233.5 F 0.0
20:00 227.6 F 0.0
25:00 226.1 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:37
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

12/21/2022 03:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____