

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1003

Name: RITTER2  
Sterilizer ID: V2313228  
04/19/2022 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
2:00 192.2 F 1.2  
4:00 209.8 F 6.8  
6:00 219.1 F 12.6  
8:00 229.7 F 9.2  
10:00 238.1 F 9.6  
12:00 243.2 F 12.6  
14:00 244.7 F 13.3  
16:00 257.7 F 20.3  
18:00 269.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.5 F 29.7  
2:00 271.6 F 29.7  
3:00 272.3 F 29.5  
4:00 271.7 F 29.8  
  
Min 271.3 F 29.1  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 211.3 F 0.0  
10:00 219.7 F 0.0  
15:00 225.5 F 0.0  
20:00 226.6 F 0.0  
25:00 228.2 F 0.0  
30:00 229.4 F 0.0  
  
Drying Complete

FILLING: 1:60  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:15  
CYCLE COMPLETE

04/19/2022 10:27 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_