

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 004536  
9:38 17/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:58  
STERILIZING 2:57  
136.5 C 222kPa 2:57  
136.3 C 224kPa 3:27  
136.4 C 223kPa 3:57  
136.4 C 219kPa 4:27  
136.2 C 223kPa 4:57  
136.4 C 221kPa 5:27  
136.4 C 223kPa 5:57  
136.1 C 222kPa 6:27  
136.1 C 222kPa 6:27  
VENTING 6:27  
AIR DRYING 7:17  
CYCLE COMPLETE 7:24  
-----

Digital Signature #  
8FE4DA4A083825B9  
-----