
Midmark M11 - v1.0.0
Total Cycles: 5380

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2025 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 160.8 F 3.2
4:00 175.7 F 7.0
6:00 196.1 F 11.3
8:00 214.1 F 4.4
10:00 226.0 F 8.9
12:00 240.2 F 11.8
14:00 238.3 F 10.8
16:00 235.2 F 8.6
18:00 249.2 F 15.1
20:00 260.3 F 21.4
22:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.3 F 28.9
2:00 272.2 F 29.4
3:00 273.3 F 29.7
4:00 272.2 F 29.1

Min 271.0 F 28.4
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 199.2 F 0.0
10:00 209.8 F 0.0
15:00 211.1 F 0.0
20:00 210.3 F 0.0
25:00 217.5 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

02/28/2025 04:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____