

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3162

Name: RITTER2  
Sterilizer ID: V2313228  
11/27/2023 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.6 F 0.0  
2:00 143.7 F 1.3  
4:00 161.8 F 4.4  
6:00 186.4 F 8.8  
8:00 210.3 F 12.3  
10:00 216.2 F 8.3  
12:00 230.9 F 12.3  
14:00 236.6 F 11.0  
16:00 236.3 F 9.7  
18:00 250.7 F 16.5  
20:00 262.7 F 23.5  
22:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.7 F 30.0  
2:00 271.3 F 29.4  
3:00 272.6 F 30.1  
4:00 271.8 F 29.5

Min 271.3 F 29.3  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
5:00 220.5 F 0.0  
10:00 204.1 F 0.0  
15:00 197.4 F 0.0  
20:00 194.3 F 0.0  
25:00 191.4 F 0.0  
30:00 189.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

11/27/2023 02:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_