
Midmark M11 - v1.0.5
Total Cycles: 1067

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 170.5 F 3.3
4:00 198.9 F 8.1
6:00 215.0 F 12.2
8:00 226.7 F 9.0
10:00 235.8 F 10.2
12:00 242.0 F 12.7
14:00 242.4 F 11.9
16:00 255.4 F 18.8
18:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.6 F 29.8
4:00 271.8 F 29.1

Min 271.4 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 209.9 F 0.0
10:00 218.6 F 0.0
15:00 219.8 F 0.0
20:00 218.7 F 0.0
25:00 219.9 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

05/03/2022 04:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____