

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4175

Name: RITTER2  
Sterilizer ID: V2313228  
12/03/2024 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 72.6 F 0.0  
4:00 73.5 F 0.0  
6:00 87.4 F 0.1  
8:00 136.7 F 0.2  
10:00 159.8 F 0.5  
12:00 182.1 F 0.7  
14:00 198.6 F 2.9  
16:00 216.5 F 5.7  
18:00 223.9 F 9.1  
20:00 233.0 F 12.7  
22:00 232.2 F 8.8  
24:00 242.4 F 13.5  
26:00 240.7 F 11.3  
28:00 238.3 F 9.8  
30:00 250.6 F 15.7  
32:00 260.7 F 21.4  
34:00 270.1 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.4 F 29.4  
2:00 272.4 F 29.0  
3:00 271.5 F 28.5  
4:00 271.5 F 28.9  
  
Min 271.1 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 214.9 F 0.0  
10:00 229.1 F 0.0  
15:00 228.0 F 0.0  
20:00 224.0 F 0.0  
25:00 224.0 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 35:03  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:13:41  
CYCLE COMPLETE

12/03/2024 11:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_