
Midmark M11 - v1.0.5
Total Cycles: 1042

Name: RITTER2
Sterilizer ID: V2313228
04/27/2022 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 185.9 F 3.4
4:00 188.7 F 8.1
6:00 209.6 F 13.3
8:00 225.4 F 8.7
10:00 237.3 F 11.1
12:00 240.4 F 12.1
14:00 240.9 F 11.2
16:00 253.9 F 18.1
18:00 265.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.6 F 29.6
2:00 272.5 F 29.7
3:00 271.6 F 29.2
4:00 271.3 F 29.4

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 210.5 F 0.0
10:00 222.5 F 0.0
15:00 228.0 F 0.0
20:00 233.0 F 0.0
25:00 226.9 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

04/27/2022 11:44 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____