
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001061
10:23 30/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:11
STERILIZING 4:12
136.0 C 220kPa 4:12
136.3 C 221kPa 4:42
136.4 C 224kPa 5:12
136.2 C 221kPa 5:42
136.4 C 225kPa 6:12
136.4 C 225kPa 6:42
136.1 C 221kPa 7:12
136.4 C 224kPa 7:42
136.4 C 224kPa 7:42
VENTING 7:42
AIR DRYING 8:37
CYCLE COMPLETE 8:43

Digital Signature #
3E2C4C88AA106D83
