
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002296
13:11 25/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:52
STERILIZING 2:54
136.5 C 224kPa 2:54
136.2 C 224kPa 3:24
135.9 C 219kPa 3:54
136.3 C 225kPa 4:24
136.3 C 224kPa 4:54
136.4 C 222kPa 5:24
136.3 C 224kPa 5:54
136.4 C 223kPa 6:24
136.4 C 223kPa 6:24
VENTING 6:24
AIR DRYING 7:18
CYCLE COMPLETE 14:22

Digital Signature #
161D144A6022687D
