

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5461

Name: RITTER 1  
Sterilizer ID: V2178198  
03/28/2025 11:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.0 F 0.0  
2:00 140.5 F 2.0  
4:00 148.9 F 4.9  
6:00 175.3 F 8.5  
8:00 195.8 F 12.7  
10:00 208.1 F 4.8  
12:00 222.4 F 9.5  
14:00 235.1 F 9.3  
16:00 238.6 F 11.0  
18:00 235.7 F 8.8  
20:00 249.5 F 15.2  
22:00 260.8 F 21.5  
24:00 271.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.1 F 29.1  
2:00 271.1 F 28.4  
3:00 272.1 F 29.2  
4:00 271.5 F 28.6

Min 271.1 F 28.3  
Max 272.2 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 203.5 F 0.0  
10:00 206.7 F 0.0  
15:00 215.7 F 0.0  
20:00 222.3 F 0.0  
25:00 226.9 F 0.0  
30:00 230.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:40  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:03:17  
CYCLE COMPLETE

03/28/2025 01:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_