
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001177
15:26 20/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:22
STERILIZING 3:30
136.0 C 220kPa 3:30
136.4 C 223kPa 4:00
136.3 C 223kPa 4:30
136.4 C 224kPa 5:00
136.0 C 221kPa 5:30
136.1 C 221kPa 6:00
136.2 C 222kPa 6:30
136.4 C 225kPa 7:00
136.4 C 225kPa 7:00
VENTING 7:00
AIR DRYING 7:52
CYCLE COMPLETE 22:52

Digital Signature #
542C34DD38764A23
