
Midmark M11 - v1.0.0
Total Cycles: 3410

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 07:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 79.5 F 0.0
4:00 87.0 F 0.0
6:00 115.4 F 0.0
8:00 152.6 F 0.7
10:00 181.7 F 1.0
12:00 203.2 F 3.2
14:00 214.6 F 5.8
16:00 223.6 F 8.9
18:00 233.2 F 12.5
20:00 232.7 F 8.3
22:00 244.0 F 13.3
24:00 241.4 F 11.0
26:00 237.8 F 9.2
28:00 251.3 F 15.6
30:00 262.5 F 22.0
32:00 272.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.4 F 28.4
2:00 271.8 F 28.3
3:00 272.1 F 28.2
4:00 272.0 F 28.5

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 205.6 F 0.0
10:00 214.4 F 0.0
15:00 216.7 F 0.0
20:00 220.7 F 0.0
25:00 227.1 F 0.0
30:00 225.5 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

06/20/2023 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____