
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000199
9:12 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:19
STERILIZING 4:51
136.0 C 222kPa 4:51
136.2 C 231kPa 5:21
136.4 C 231kPa 5:51
136.4 C 231kPa 6:21
136.4 C 233kPa 6:51
136.4 C 233kPa 7:21
136.3 C 231kPa 7:51
136.4 C 233kPa 8:21
136.4 C 233kPa 8:21
VENTING 8:21
AIR DRYING 9:23
CYCLE COMPLETE 9:26

Digital Signature #
2A126B6A88C659AD
