
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007920
7:37 03/10/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:03
STERILIZING 4:22
136.4 C 222kPa 4:22
136.3 C 223kPa 4:52
136.0 C 222kPa 5:22
136.1 C 222kPa 5:52
136.2 C 220kPa 6:22
136.4 C 223kPa 6:52
136.3 C 219kPa 7:22
136.4 C 221kPa 7:52
136.4 C 221kPa 7:52
VENTING 7:52
AIR DRYING 8:43
CYCLE COMPLETE 38:44

Digital Signature #
02497289CA2924A4
