
Midmark M11 - v1.0.5
Total Cycles: 736

Name: RITTER2
Sterilizer ID: V2313228
01/31/2022 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 181.1 F 3.3
4:00 205.8 F 9.5
6:00 224.1 F 6.4
8:00 239.2 F 13.0
10:00 241.9 F 11.8
12:00 242.8 F 11.7
14:00 257.4 F 19.4
16:00 269.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.8 F 28.3
2:00 272.3 F 28.7
3:00 272.8 F 29.1
4:00 271.9 F 28.8

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 226.5 F 0.0
10:00 225.2 F 0.0
15:00 228.6 F 0.0
20:00 228.2 F 0.0
25:00 230.5 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 16:49
STERILIZING: 4:00
VENTING: 2:50
DRYING: 30:00
TOTAL CYCLE: 00:55:52
CYCLE COMPLETE

01/31/2022 12:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____