
Midmark M11 - v1.0.0
Total Cycles: 4202

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2024 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 76.2 F 0.3
4:00 90.5 F 1.3
6:00 121.9 F 3.6
8:00 147.4 F 6.4
10:00 170.3 F 9.9
12:00 190.9 F 13.3
14:00 211.3 F 4.0
16:00 221.9 F 7.0
18:00 229.0 F 10.4
20:00 232.7 F 8.9
22:00 237.2 F 9.8
24:00 239.0 F 9.8
26:00 245.1 F 12.6
28:00 256.3 F 18.5
30:00 266.3 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.1
3:00 272.6 F 28.7
4:00 271.4 F 28.0

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 237.3 F 0.0
10:00 230.1 F 0.0
15:00 229.2 F 0.0
20:00 230.2 F 0.0
25:00 224.7 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

01/15/2024 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____