

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4218

Name: RITTER2  
Sterilizer ID: V2313228  
12/16/2024 01:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.8 F 0.0  
2:00 120.0 F 0.5  
4:00 130.2 F 2.9  
6:00 153.1 F 6.3  
8:00 175.8 F 10.2  
10:00 200.0 F 7.6  
12:00 216.6 F 7.0  
14:00 227.4 F 11.3  
16:00 228.9 F 7.8  
18:00 241.5 F 13.2  
20:00 241.3 F 11.6  
22:00 253.1 F 17.5  
24:00 263.5 F 23.8  
26:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.1  
1:00 272.4 F 30.0  
2:00 272.4 F 29.5  
3:00 271.2 F 29.0  
4:00 271.7 F 29.8

Min 271.0 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 217.2 F 0.0  
10:00 229.2 F 0.0  
15:00 222.7 F 0.0  
20:00 206.2 F 0.0  
25:00 210.2 F 0.0  
30:00 211.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:20  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:07  
CYCLE COMPLETE

12/16/2024 02:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_