

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1721

Name: RITTER 1  
Sterilizer ID: V2178198  
03/16/2022 08:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.3 F 0.0  
2:00 126.2 F 0.1  
4:00 176.0 F 1.6  
6:00 188.6 F 3.8  
8:00 206.6 F 7.3  
10:00 224.5 F 11.5  
12:00 223.7 F 5.5  
14:00 235.5 F 10.2  
16:00 235.4 F 8.2  
18:00 243.9 F 12.5  
20:00 241.2 F 10.6  
22:00 253.6 F 16.9  
24:00 264.6 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.2  
1:00 272.2 F 28.8  
2:00 271.5 F 28.5  
3:00 271.7 F 28.3  
4:00 272.3 F 28.9

Min 271.2 F 28.1  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 216.6 F 0.0  
10:00 237.0 F 0.0  
15:00 233.9 F 0.0  
20:00 220.4 F 0.0  
25:00 220.1 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:47  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:03:50  
CYCLE COMPLETE

03/16/2022 10:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_