
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000855
16:59 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 3:01
STERILIZING 4:07
136.0 C 229kPa 4:07
136.3 C 233kPa 4:37
136.4 C 233kPa 5:07
136.4 C 233kPa 5:37
136.4 C 232kPa 6:07
136.4 C 233kPa 6:37
136.3 C 232kPa 7:07
136.4 C 233kPa 7:37
136.4 C 233kPa 7:37
VENTING 7:37
AIR DRYING 8:31
CYCLE COMPLETE 23:31

Digital Signature #
24BEC9A4F0F8F75A
