
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000165
6:57 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:37
PRESSURIZING 4:37
STERILIZING 6:01
136.0 C 220kPa 6:01
136.0 C 222kPa 6:31
136.2 C 223kPa 7:01
136.2 C 221kPa 7:31
136.4 C 224kPa 8:01
136.4 C 223kPa 8:31
136.3 C 224kPa 9:01
136.0 C 222kPa 9:31
136.0 C 222kPa 9:31
VENTING 9:31
AIR DRYING 10:30
CYCLE COMPLETE 25:30

Digital Signature #
F1FA1EB77FAFCBF6
