

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1321

Name: RITTER2  
Sterilizer ID: V2313228  
07/20/2022 07:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.8 F 0.0  
2:00 81.0 F 0.0  
4:00 132.8 F 0.0  
6:00 174.3 F 0.0  
8:00 150.9 F 0.5  
10:00 180.7 F 0.9  
12:00 207.3 F 3.0  
14:00 217.5 F 6.1  
16:00 229.3 F 9.4  
18:00 238.6 F 13.2  
20:00 233.4 F 9.0  
22:00 242.5 F 12.1  
24:00 240.7 F 11.7  
26:00 238.7 F 10.2  
28:00 251.1 F 16.4  
30:00 262.0 F 22.6  
32:00 272.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.5 F 29.5  
2:00 271.4 F 28.9  
3:00 272.5 F 29.5  
4:00 271.5 F 28.8  
  
Min 271.3 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 225.3 F 0.0  
10:00 232.5 F 0.0  
15:00 222.9 F 0.0  
20:00 219.6 F 0.0  
25:00 221.8 F 0.0  
30:00 223.1 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:31  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:10:28  
CYCLE COMPLETE

07/20/2022 08:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_